

# Underwriters Laboratories (UL LLC) Safety Certification (Manufacturing Factory) Report



Model: LFM200MXXXB and LFM200MXXXC  
"XXX" can be 120, 150, 240, 280, 300, 360, 480 or 540

Device Description: Switching Power Supply

Applicant: CINCON ELECTRONICS CO LTD  
8-1 FU KUNG RD, FU HSING PARK, FU HSING HSIANG,  
CHANGHUA HSIEN, 506, TAIWAN

Manufacturer: Same as Applicant

  

Manufacturing Facility(ies): DONGGUAN CINCON ELECTRONICS LTD  
66 YUANYING RD LIAOBU TOWN, DONGGUAN,  
GUANGDONG 523412 CHINA

CINCON ELECTRONICS CO LTD  
8-1 FU KUNG RD, FU HSING PARK, FU HSING HSIANG,  
CHANGHUA HSIEN, 506, TAIWAN

Report No.: E252331-D1030-1/A0/C0(M)

Report (Re)Issue Date: 2024-04-xx

Base Standard(s): ANSI/AAMI ES60601-1:2005 & A1:2012 & A2:2021 (consolidated version) –  
revision date: 2022/02/01  
CAN/CSA-C22.2 No. 60601-1:14 (IEC 60601-1: 2005+A1:2012+A2:2020,  
MOD) (consolidated version) – revision date: 2022/July

Additional Standards: N/A

Report Types: This report consists of the following report types:  
[ Yes ] US Certification (UL Recognition)  
[ Yes ] CAN Certification (cUL Recognition)

This report covers the Safety evaluation of the referenced model(s) according to the standard(s) specified above.

This is the Manufacturing Factory report only, which is used as part of the factory FUS inspections.